

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020525\
 Data File : VN085684.D
 Acq On : 05 Feb 2025 17:59
 Operator : JC\MD
 Sample : Q1287-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BAY-DRUMS-B

Quant Time: Feb 06 00:45:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	197178	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	389563	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	350734	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	142442	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	153675	48.282	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.560%		
35) Dibromofluoromethane	8.165	113	127860	47.310	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.620%		
50) Toluene-d8	10.565	98	487912	50.812	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	101.620%		
62) 4-Bromofluorobenzene	12.847	95	149457	45.501	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	91.000%		
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	61480	59.782	ug/l	96
18) Methyl Acetate	5.030	43	3327	1.064	ug/l #	77
20) Methylene Chloride	5.265	84	3208	1.260	ug/l #	71
43) Isopropyl Acetate	8.682	43	163091	26.585	ug/l #	87
86) p-Isopropyltoluene	13.723	119	11780	1.629	ug/l	99
95) Naphthalene	15.635	128	44271	6.918	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020525\
Data File : VN085684.D
Acq On : 05 Feb 2025 17:59
Operator : JC\MD
Sample : Q1287-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BAY-DRUMS-B

Quant Time: Feb 06 00:45:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

