

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020525\
 Data File : VN085687.D
 Acq On : 05 Feb 2025 19:11
 Operator : JC\MD
 Sample : Q1287-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BAY-DRUMS-E

Quant Time: Feb 06 00:46:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	205540	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	405236	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	373093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	143504	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	160628	48.413	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.820%
35) Dibromofluoromethane	8.165	113	135146	48.072	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.140%
50) Toluene-d8	10.565	98	514058	51.464	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.920%
62) 4-Bromofluorobenzene	12.847	95	151046	44.206	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.420%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	46544	43.417	ug/l	89
20) Methylene Chloride	5.283	84	3085	1.162	ug/l #	82
43) Isopropyl Acetate	8.683	43	163051	25.551	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020525\
Data File : VN085687.D
Acq On : 05 Feb 2025 19:11
Operator : JC\MD
Sample : Q1287-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BAY-DRUMS-E

Quant Time: Feb 06 00:46:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

