

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020525\
 Data File : VN085688.D
 Acq On : 05 Feb 2025 19:35
 Operator : JC\MD
 Sample : Q1289-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FL-DRUMS-A

Quant Time: Feb 06 00:47:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	206383	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	404979	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	373277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	142618	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	160819	48.273	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.540%
35) Dibromofluoromethane	8.165	113	135100	48.086	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.180%
50) Toluene-d8	10.565	98	516501	51.742	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.480%
62) 4-Bromofluorobenzene	12.847	95	154152	45.144	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.280%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	36860	34.243	ug/l	96
20) Methylene Chloride	5.265	84	3130	1.175	ug/l	92
43) Isopropyl Acetate	8.688	43	164831	25.846	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020525\
Data File : VN085688.D
Acq On : 05 Feb 2025 19:35
Operator : JC\MD
Sample : Q1289-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FL-DRUMS-A

Quant Time: Feb 06 00:47:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

