

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020619\
 Data File : VN053633.D
 Acq On : 6 Feb 2019 14:30
 Operator : JC/SP
 Sample : PB116860ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116860ZHE#06

Quant Time: Feb 07 04:25:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	748651	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1245665	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1150887	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	482418	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	395261	52.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.26%	
35) Dibromofluoromethane	7.59	113	389814	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
50) Toluene-d8	10.09	98	1502821	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
62) 4-Bromofluorobenzene	12.40	95	505710	51.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.62%	
Target Compounds						
16) Acetone	3.82	43	25695	13.338	ug/l	95
20) Methylene Chloride	4.55	84	17817	2.204	ug/l	94
43) Isopropyl Acetate	8.17	43	29187	2.410	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020619\
Data File : VN053633.D
Acq On : 6 Feb 2019 14:30
Operator : JC/SP
Sample : PB116860ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116860ZHE#06

Quant Time: Feb 07 04:25:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

