

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020619\
 Data File : VN053634.D
 Acq On : 6 Feb 2019 14:56
 Operator : JC/SP
 Sample : PB116860ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116860ZHE#07

Quant Time: Feb 07 04:26:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	746454	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1248738	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1164095	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	511682	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	402535	53.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.50%	
35) Dibromofluoromethane	7.59	113	396851	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
50) Toluene-d8	10.09	98	1514272	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
62) 4-Bromofluorobenzene	12.40	95	521062	52.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.46%	
Target Compounds						
16) Acetone	3.82	43	27494	14.314	ug/l	100
20) Methylene Chloride	4.55	84	19852	2.463	ug/l	94
25) 2-Butanone	6.84	43	5072	1.851	ug/l	94
43) Isopropyl Acetate	8.17	43	38073	3.136	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020619\
Data File : VN053634.D
Acq On : 6 Feb 2019 14:56
Operator : JC/SP
Sample : PB116860ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116860ZHE#07

Quant Time: Feb 07 04:26:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

