

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020619\
 Data File : VN053640.D
 Acq On : 6 Feb 2019 17:37
 Operator : JC/SP
 Sample : K1356-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-01-020519-A

Quant Time: Feb 07 04:51:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	775236	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1263839	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1149330	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	498655	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	404257	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
35) Dibromofluoromethane	7.59	113	397080	49.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.44%	
50) Toluene-d8	10.09	98	1518639	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
62) 4-Bromofluorobenzene	12.40	95	507236	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
Target Compounds						
16) Acetone	3.82	43	24946	12.505	ug/l	96
20) Methylene Chloride	4.55	84	2040489	243.809	ug/l	98
43) Isopropyl Acetate	8.17	43	39821	3.241	ug/l #	79
95) Naphthalene	15.13	128	47837	3.844	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020619\
Data File : VN053640.D
Acq On : 6 Feb 2019 17:37
Operator : JC/SP
Sample : K1356-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SU-01-020519-A

Quant Time: Feb 07 04:51:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

