

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020619\
 Data File : VN053642.D
 Acq On : 6 Feb 2019 18:30
 Operator : JC/SP
 Sample : K1358-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-05-020519

Quant Time: Feb 07 05:56:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	798424	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1293823	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1147369	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	470414	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	406606	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
35) Dibromofluoromethane	7.59	113	399877	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
50) Toluene-d8	10.09	98	1543566	50.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.56%	
62) 4-Bromofluorobenzene	12.40	95	501832	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
Target Compounds						
16) Acetone	3.82	43	17881	8.703	ug/l	96
20) Methylene Chloride	4.55	84	2512212	291.456	ug/l	98
43) Isopropyl Acetate	8.17	43	40194	3.195	ug/l #	82
95) Naphthalene	15.13	128	11152	2.257	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020619\
Data File : VN053642.D
Acq On : 6 Feb 2019 18:30
Operator : JC/SP
Sample : K1358-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TR-05-020519

Quant Time: Feb 07 05:56:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

