

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020623\
 Data File : VN076549.D
 Acq On : 06 Feb 2023 18:12
 Operator : JC\MD
 Sample : 01407-01
 Misc : 1.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30984

Quant Time: Feb 07 05:48:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 07 01:44:01 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	509833	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	823203	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	685994	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	287933	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	279042	45.967	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.940%
35) Dibromofluoromethane	8.171	113	242287	47.561	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.120%
50) Toluene-d8	10.565	98	934532	48.284	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.560%
62) 4-Bromofluorobenzene	12.853	95	298008	44.424	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020623\
Data File : VN076549.D
Acq On : 06 Feb 2023 18:12
Operator : JC\MD
Sample : 01407-01
Misc : 1.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30984

Quant Time: Feb 07 05:48:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020623W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 07 01:44:01 2023
Response via : Initial Calibration

