

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020624\
 Data File : VN080925.D
 Acq On : 06 Feb 2024 18:33
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Feb 07 05:33:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 07 05:25:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	200000	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	441062	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	529427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	235831	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	183965	72.915	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	145.820%#
35) Dibromofluoromethane	8.165	113	168424	71.163	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	142.320%#
50) Toluene-d8	10.571	98	719450	84.041	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	168.080%#
62) 4-Bromofluorobenzene	12.853	95	258850	83.234	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	166.460%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020624\
Data File : VN080925.D
Acq On : 06 Feb 2024 18:33
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Feb 07 05:33:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 07 05:25:05 2024
Response via : Initial Calibration

