


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020624\
Data File : VN080917.D
Acq On : 06 Feb 2024 13:39
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
Title : SW846 8260
Last Update : Wed Feb 07 06:22:23 2024



AutoFind: Scans 1852, 1853, 1854; Background Corrected with Scan 1843

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.1	15139	PASS
75	95	30	60	55.3	41725	PASS
95	95	100	100	100.0	75408	PASS
96	95	5	9	6.6	4973	PASS
173	174	0.00	2	1.3	709	PASS
174	95	50	100	73.9	55701	PASS
175	174	5	9	8.5	4718	PASS
176	174	95	101	95.3	53107	PASS
177	176	5	9	7.4	3945	PASS

Average of 12.847 to 12.859 min.: VN080917.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
36.00	639	48.95	3448	63.00	3304	72.90	1377
37.00	4330	50.00	15139	64.05	354	73.05	189
38.00	3941	51.00	4526	64.70	88	73.95	1431
38.95	1026	51.95	263	65.15	287	75.00	4172
40.00	467	54.90	53	65.90	61	76.00	3429
42.90	60	55.95	1488	66.85	348	76.70	129
43.95	827	57.05	2619	67.95	7396	76.85	378
44.95	765	57.80	84	68.95	7633	77.30	165
46.10	68	59.95	840	69.95	898	77.65	216
46.90	510	60.95	4372	71.90	144	78.10	62
47.80	472	62.00	4106	72.05	472	78.90	3897

Average of 12.847 to 12.859 min.: VN080917.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
79.85	962	91.95	3200	105.85	471	129.95	353
80.85	3786	92.95	4093	111.90	56	130.70	93
81.85	986	93.95	10769	115.80	346	133.60	66
82.90	61	95.00	75408	116.80	605	135.00	211
83.10	56	95.95	4973	117.60	107	136.85	220
85.70	76	96.70	53	117.85	309	140.60	322
86.90	2991	97.00	174	118.70	104	140.85	676
87.80	1087	103.80	339	118.90	282	142.80	881
88.00	2119	104.10	180	127.65	253	144.75	128
90.70	203	104.40	50	128.00	134	145.85	151
91.05	251	105.00	98	128.80	59	147.60	137

Average of 12.847 to 12.859 min.: VN080917.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
154.60	106	176.90	3945				
155.20	105	206.80	68				
158.80	66						
160.60	53						
171.60	176						
171.75	362						
172.60	175						
172.90	709						
173.90	55701						
174.85	4718						
175.90	53107						