

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020719\
 Data File : VN053647.D
 Acq On : 7 Feb 2019 11:17
 Operator : JC/SP
 Sample : VN0207WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0207WBL01

Quant Time: Feb 09 00:46:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	712537	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1164856	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1071237	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	427306	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	375139	52.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.96%	
35) Dibromofluoromethane	7.59	113	373591	50.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.52%	
50) Toluene-d8	10.09	98	1404018	50.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.60%	
62) 4-Bromofluorobenzene	12.40	95	459725	49.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.76%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020719\
Data File : VN053647.D
Acq On : 7 Feb 2019 11:17
Operator : JC/SP
Sample : VN0207WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0207WBL01

Quant Time: Feb 09 00:46:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

