

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020719\
 Data File : VN053649.D
 Acq On : 7 Feb 2019 12:19
 Operator : JC/SP
 Sample : PB116887TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116887TB

Quant Time: Feb 09 00:49:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	765693	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1277445	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1163841	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	470501	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	410445	53.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.86%	
35) Dibromofluoromethane	7.59	113	407531	50.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.98%	
50) Toluene-d8	10.09	98	1547343	51.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.10%	
62) 4-Bromofluorobenzene	12.40	95	503038	49.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.54%	
Target Compounds						
16) Acetone	3.82	43	14605	7.413	ug/l	95
20) Methylene Chloride	4.56	84	9916	1.200	ug/l	89
25) 2-Butanone	6.85	43	4835	1.721	ug/l #	85
43) Isopropyl Acetate	8.17	43	31195	2.512	ug/l #	89
95) Naphthalene	15.14	128	5099	1.974	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020719\
Data File : VN053649.D
Acq On : 7 Feb 2019 12:19
Operator : JC/SP
Sample : PB116887TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB116887TB

Quant Time: Feb 09 00:49:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

