

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020719\
 Data File : VN053650.D
 Acq On : 7 Feb 2019 12:46
 Operator : JC/SP
 Sample : PB116887ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116887ZHE#01

Quant Time: Feb 09 00:50:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	772136	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1280277	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1155562	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	479193	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	411131	53.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.16%	
35) Dibromofluoromethane	7.59	113	407806	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
50) Toluene-d8	10.09	98	1536458	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
62) 4-Bromofluorobenzene	12.40	95	507190	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
Target Compounds						
16) Acetone	3.82	43	181870	91.536	ug/l	99
20) Methylene Chloride	4.55	84	21782	2.613	ug/l	91
25) 2-Butanone	6.85	43	4880	1.722	ug/l #	89
43) Isopropyl Acetate	8.17	43	34263	2.753	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020719\
Data File : VN053650.D
Acq On : 7 Feb 2019 12:46
Operator : JC/SP
Sample : PB116887ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB116887ZHE#01

Quant Time: Feb 09 00:50:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

