

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020719\
 Data File : VN053656.D
 Acq On : 7 Feb 2019 15:26
 Operator : JC/SP
 Sample : PB116887ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116887ZHE#07

Quant Time: Feb 09 00:59:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	749383	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1237491	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1140554	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	467817	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	405109	53.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.78%	
35) Dibromofluoromethane	7.59	113	398313	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
50) Toluene-d8	10.09	98	1494793	50.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.82%	
62) 4-Bromofluorobenzene	12.40	95	492000	50.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.48%	
Target Compounds						
16) Acetone	3.82	43	166718	86.457	ug/l	99
20) Methylene Chloride	4.55	84	20929	2.587	ug/l	96
25) 2-Butanone	6.84	43	4253	1.546	ug/l	100
43) Isopropyl Acetate	8.18	43	36693	3.050	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020719\
Data File : VN053656.D
Acq On : 7 Feb 2019 15:26
Operator : JC/SP
Sample : PB116887ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB116887ZHE#07

Quant Time: Feb 09 00:59:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

