

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020719\
 Data File : VN053666.D
 Acq On : 7 Feb 2019 19:54
 Operator : JC/SP
 Sample : K1389-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 M-4

Manual Integrations
 APPROVED

MMDadoda
 2/11/2019 12:21:56 PM

Quant Time: Feb 09 01:42:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	762025	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1240989	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1136298	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	467765	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	397713	52.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.06%	
35) Dibromofluoromethane	7.59	113	394364	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
50) Toluene-d8	10.09	98	1489514	50.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.18%	
62) 4-Bromofluorobenzene	12.40	95	491603	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
Target Compounds						
16) Acetone	3.82	43	15624	7.968	ug/l	94
20) Methylene Chloride	4.52	84	45559102m	5538.037	ug/l	
25) 2-Butanone	6.84	43	6246	2.233	ug/l	91
43) Isopropyl Acetate	8.17	43	42883	3.554	ug/l #	76
95) Naphthalene	15.13	128	9866	2.200	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020719\
Data File : VN053666.D
Acq On : 7 Feb 2019 19:54
Operator : JC/SP
Sample : K1389-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
M-4

Manual Integrations
APPROVED

MMDadoda
2/11/2019 12:21:56 PM

Quant Time: Feb 09 01:42:07 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

