

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020719\
 Data File : VN053668.D
 Acq On : 7 Feb 2019 20:47
 Operator : JC/SP
 Sample : K1389-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 M-5S

Manual Integrations
 APPROVED

MMDadoda
 2/11/2019 12:25:19 PM

Quant Time: Feb 09 01:49:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	753651	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1227680	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1105500	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	443236	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	396982	52.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.02%	
35) Dibromofluoromethane	7.59	113	388989	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
50) Toluene-d8	10.09	98	1466233	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
62) 4-Bromofluorobenzene	12.40	95	480982	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
Target Compounds						
16) Acetone	3.82	43	14318	7.383	ug/l	96
20) Methylene Chloride	4.52	84	42937811m	5277.394	ug/l	
25) 2-Butanone	6.85	43	5712	2.065	ug/l #	87
43) Isopropyl Acetate	8.18	43	35663	2.988	ug/l #	80
95) Naphthalene	15.13	128	6505	2.059	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020719\
Data File : VN053668.D
Acq On : 7 Feb 2019 20:47
Operator : JC/SP
Sample : K1389-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
M-5S

Manual Integrations
APPROVED

MMDadoda
2/11/2019 12:25:19 PM

Quant Time: Feb 09 01:49:22 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

