

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN020720\
 Data File : VN060017.D
 Acq On : 7 Feb 2020 8:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 07 23:30:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.534	0.480	10.1	77	0.00
3 P	Chloromethane	0.572	0.497	13.1	83	0.00
4 C	Vinyl Chloride	0.655	0.646	1.4#	88	0.00
5 T	Bromomethane	0.400	0.380	5.0	89	0.00
6 T	Chloroethane	0.310	0.311	-0.3	96	0.00
7 T	Trichlorofluoromethane	0.787	0.780	0.9	93	0.00
8 T	Diethyl Ether	0.276	0.286	-3.6	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.438	0.464	-5.9	98	0.00
10 T	Methyl Iodide	0.708	0.596	15.8	79	0.00
11 T	Tert butyl alcohol	0.093	0.093	0.0	93	0.00
12 CM	1,1-Dichloroethene	0.457	0.430	5.9#	90	0.00
13 T	Acrolein	0.066	0.049	25.8#	73	0.00
14 T	Allyl chloride	0.617	0.647	-4.9	99	0.00
15 T	Acrylonitrile	0.213	0.241	-13.1	103	0.00
16 T	Acetone	0.224	0.286	-27.7#	139	0.00
17 T	Carbon Disulfide	1.303	1.011	22.4	74	0.00
18 T	Methyl Acetate	0.649	0.742	-14.3	108	0.00
19 T	Methyl tert-butyl Ether	1.491	1.629	-9.3	104	0.00
20 T	Methylene Chloride	0.557	0.516	7.4	99	0.00
21 T	trans-1,2-Dichloroethene	0.486	0.478	1.6	94	0.00
22 T	Diisopropyl ether	1.397	1.576	-12.8	107	0.00
23 T	Vinyl Acetate	1.082	1.200	-10.9	111	0.00
24 P	1,1-Dichloroethane	0.854	0.912	-6.8	102	0.00
25 T	2-Butanone	0.314	0.368	-17.2	114	0.00
26 T	2,2-Dichloropropane	0.821	0.878	-6.9	104	0.00
27 T	cis-1,2-Dichloroethene	0.559	0.591	-5.7	100	0.00
28 T	Bromochloromethane	0.341	0.350	-2.6	99	0.00
29 T	Tetrahydrofuran	0.195	0.209	-7.2	102	0.00
30 C	Chloroform	0.885	0.987	-11.5#	106	0.00
31 T	Cyclohexane	0.822	0.748	9.0	90	0.00
32 T	1,1,1-Trichloroethane	0.818	0.890	-8.8	102	0.00
33 S	1,2-Dichloroethane-d4	0.555	0.585	-5.4	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.298	0.310	-4.0	96	0.00
36 T	1,1-Dichloropropene	0.435	0.442	-1.6	96	0.00
37 T	Ethyl Acetate	0.408	0.431	-5.6	104	0.00
38 T	Carbon Tetrachloride	0.474	0.497	-4.9	101	0.00
39 T	Methylcyclohexane	0.537	0.530	1.3	94	0.00
40 TM	Benzene	1.299	1.343	-3.4	100	0.00
41 T	Methacrylonitrile	0.201	0.215	-7.0	103	0.00
42 TM	1,2-Dichloroethane	0.440	0.498	-13.2	108	0.00
43 T	Isopropyl Acetate	0.702	0.753	-7.3	103	0.00
44 TM	Trichloroethene	0.354	0.375	-5.9	98	0.00
45 C	1,2-Dichloropropane	0.321	0.360	-12.1#	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN020720\
 Data File : VN060017.D
 Acq On : 7 Feb 2020 8:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 07 23:30:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.227	0.247	-8.8	103	0.00
47 T	Bromodichloromethane	0.457	0.516	-12.9	107	0.00
48 T	Methyl methacrylate	0.314	0.332	-5.7	100	0.00
49 T	1,4-Dioxane	0.006	0.006	0.0	94	0.00
50 S	Toluene-d8	1.155	1.164	-0.8	94	0.00
51 T	4-Methyl-2-Pentanone	0.406	0.454	-11.8	106	0.00
52 CM	Toluene	0.820	0.879	-7.2#	101	0.00
53 T	t-1,3-Dichloropropene	0.510	0.583	-14.3	106	0.00
54 T	cis-1,3-Dichloropropene	0.537	0.604	-12.5	104	0.00
55 T	1,1,2-Trichloroethane	0.325	0.363	-11.7	108	0.00
56 T	Ethyl methacrylate	0.496	0.539	-8.7	102	0.00
57 T	1,3-Dichloropropane	0.516	0.588	-14.0	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.150	0.185	-23.3	105	0.00
59 T	2-Hexanone	0.299	0.362	-21.1	109	0.00
60 T	Dibromochloromethane	0.370	0.426	-15.1	107	0.00
61 T	1,2-Dibromoethane	0.340	0.381	-12.1	104	0.00
62 S	4-Bromofluorobenzene	0.416	0.448	-7.7	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
64 T	Tetrachloroethene	0.408	0.397	2.7	89	0.00
65 PM	Chlorobenzene	0.983	1.048	-6.6	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.413	-8.7	108	0.00
67 C	Ethyl Benzene	1.784	1.881	-5.4#	105	0.00
68 T	m/p-Xylenes	0.674	0.706	-4.7	103	0.00
69 T	o-Xylene	0.660	0.688	-4.2	104	0.00
70 T	Styrene	1.104	1.163	-5.3	101	0.00
71 P	Bromoform	0.304	0.344	-13.2	108	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.667	3.801	-3.7	107	0.00
74 T	N-amyl acetate	1.423	1.453	-2.1	102	0.00
75 P	1,1,2,2-Tetrachloroethane	1.063	1.137	-7.0	110	0.00
76 T	1,2,3-Trichloropropane	0.841	0.939	-11.7	111	0.00
77 T	Bromobenzene	0.929	0.959	-3.2	105	0.00
78 T	n-propylbenzene	4.116	4.350	-5.7	108	0.00
79 T	2-Chlorotoluene	2.448	2.534	-3.5	108	0.00
80 T	1,3,5-Trimethylbenzene	3.084	3.200	-3.8	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.392	0.427	-8.9	109	0.00
82 T	4-Chlorotoluene	2.456	2.639	-7.5	109	0.00
83 T	tert-Butylbenzene	2.667	2.767	-3.7	107	0.00
84 T	1,2,4-Trimethylbenzene	3.051	3.221	-5.6	108	0.00
85 T	sec-Butylbenzene	3.504	3.768	-7.5	109	0.00
86 T	p-Isopropyltoluene	3.243	3.481	-7.3	111	0.00
87 T	1,3-Dichlorobenzene	1.613	1.766	-9.5	111	0.00
88 T	1,4-Dichlorobenzene	1.609	1.734	-7.8	110	0.00
89 T	n-Butylbenzene	2.765	3.113	-12.6	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN020720\
Data File : VN060017.D
Acq On : 7 Feb 2020 8:59
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 07 23:30:43 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.607	0.662	-9.1	111	0.00
91 T	1,2-Dichlorobenzene	1.595	1.702	-6.7	109	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.239	7.0	105	0.00
93 T	1,2,4-Trichlorobenzene	0.958	1.054	-10.0	105	0.00
94 T	Hexachlorobutadiene	0.528	0.576	-9.1	111	0.00
95 T	Naphthalene	2.726	2.815	-3.3	98	0.00
96 T	1,2,3-Trichlorobenzene	0.933	1.017	-9.0	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6