

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020720\
 Data File : VN060029.D
 Acq On : 7 Feb 2020 14:15
 Operator : JC/MD
 Sample : PB126640ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB126640ZHE#03

Quant Time: Feb 07 23:51:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	189222	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	294748	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	280707	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	127372	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	109768	52.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.48%	
35) Dibromofluoromethane	7.57	113	89558	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
50) Toluene-d8	10.08	98	358278	52.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.22%	
62) 4-Bromofluorobenzene	12.40	95	130642	53.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.42%	
Target Compounds						
20) Methylene Chloride	4.52	84	4839	1.284	ug/l #	85
43) Isopropyl Acetate	8.14	43	47005	11.352	ug/l #	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020720\
Data File : VN060029.D
Acq On : 7 Feb 2020 14:15
Operator : JC/MD
Sample : PB126640ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB126640ZHE#03

Quant Time: Feb 07 23:51:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1826

Delta R.T. 0.00 min

Lab File: VN060029.D

Acq: 7 Feb 2020 14:15

Instrument :

MSVOA_N

ClientSampleId :


PB126640ZHE#03

Tot Ion:168 Resp: 189222

Ion Ratio Lower Upper




168 100

99 53.7 48.2 72.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN



#20

Methylene Chloride

Concen: 1.284 ug/l

RT: 4.52 min Scan# 855

Delta R.T. 0.00 min

Lab File: VN060029.D

Acq: 7 Feb 2020 14:15

Tgt Ion: 84 Resp: 4839


Ion Ratio Lower Upper

84 100

49 105.6 106.6 160.0#

51 33.7 31.9 47.9

86 62.9 50.8 76.2






Abundance Ion 83.90 (83.60 to 84.60): VN

Ion 48.90 (48.60 to 49.60): VN

Ion 50.90 (50.60 to 51.60): VN

Ion 85.90 (85.60 to 86.60): VN









#33
 1,2-Dichloroethane-d4
 Concen: 52.237 ug/l
 RT: 8.00 min Scan# 1938
 Delta R.T. 0.00 min
 Lab File: VN060029.D
 Acq: 7 Feb 2020 14:15




Instrument :
 MSVOA_N
 ClientSampleId :
 PB126640ZHE#03


Tot Ion: 65 Resp: 109768
 Ion Ratio Lower Upper
 65 100
 67 52.1 0.0 103.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.57 min Scan# 2113
 Delta R.T. 0.00 min
 Lab File: VN060029.D
 Acq: 7 Feb 2020 14:15

Tgt Ion: 114 Resp: 294748
 Ion Ratio Lower Upper
 114 100
 63 19.8 0.0 43.6
 88 15.3 0.0 31.4









#35
 Dibromofluoromethane
 Concen: 50.951 ug/l
 RT: 7.57 min Scan# 1802
 Delta R.T. 0.00 min
 Lab File: VN060029.D
 Acq: 7 Feb 2020 14:15




Instrument :
 MSVOA_N
 ClientSampleId :
 PB126640ZHE#03


Tot Ion	Ratio	Lower	Upper
113	100		
111	104.8	83.5	125.3
192	20.9	15.4	23.2



#43
 Isopropyl Acetate
 Concen: 11.352 ug/l
 RT: 8.14 min Scan# 1981
 Delta R.T. 0.00 min
 Lab File: VN060029.D
 Acq: 7 Feb 2020 14:15

Tgt Ion	Ratio	Lower	Upper
43	100		
61	19.7	20.2	30.2#
87	13.9	9.5	14.3





#50

Toluene-d8

Concen: 52.609 ug/l

RT: 10.08 min Scan# 2583

Delta R.T. 0.00 min

Lab File: VN060029.D

Acq: 7 Feb 2020 14:15

Instrument :

MSVOA_N

ClientSampleId :





PB126640ZHE#03

Tot Ion: 98 Resp: 358278

Ion Ratio Lower Upper

98 100

100 64.8 51.2 76.8



#62

4-Bromofluorobenzene

Concen: 53.213 ug/l

RT: 12.40 min Scan# 3304

Delta R.T. 0.00 min

Lab File: VN060029.D

Acq: 7 Feb 2020 14:15




Tgt Ion: 95 Resp: 130642


Ion Ratio Lower Upper

95 100

174 85.1 0.0 151.4

176 82.8 0.0 146.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.40 min Scan# 2994
Delta R.T. 0.00 min
Lab File: VN060029.D
Acq: 7 Feb 2020 14:15

Instrument :
MSVOA_N
ClientSampleId :
PB126640ZHE#03

Tot Ion:117 Resp: 280707
Ion Ratio Lower Upper
117 100
82 53.4 45.9 68.9
119 32.3 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3597
Delta R.T. 0.00 min
Lab File: VN060029.D
Acq: 7 Feb 2020 14:15

Tgt Ion:152 Resp: 127372
Ion Ratio Lower Upper
152 100
115 57.7 30.3 90.9
150 158.2 0.0 345.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

