

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020723\
 Data File : VN076564.D
 Acq On : 07 Feb 2023 14:07
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Feb 08 01:20:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 01:16:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	644720	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	1067430	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	892073	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	323589	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	358723	47.141	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.280%
35) Dibromofluoromethane	8.171	113	326704	49.236	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.480%
50) Toluene-d8	10.570	98	1188172	47.598	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.200%
62) 4-Bromofluorobenzene	12.853	95	385459	43.831	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020723\
Data File : VN076564.D
Acq On : 07 Feb 2023 14:07
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Feb 08 01:20:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 01:16:19 2023
Response via : Initial Calibration

