

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020724\
 Data File : VN080935.D
 Acq On : 07 Feb 2024 14:27
 Operator : JC\MD
 Sample : P1362-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2

Quant Time: Feb 08 00:25:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 07 06:22:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	460712	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	931692	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	1055888	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	464008	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	269830	46.427	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.860%
35) Dibromofluoromethane	8.171	113	268277	53.661	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.320%
50) Toluene-d8	10.571	98	995436	55.047	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.100%
62) 4-Bromofluorobenzene	12.853	95	418045	63.636	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	127.280%#
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	34457	18.020	ug/l #	81
20) Methylene Chloride	5.277	84	24190	4.176	ug/l #	76
37) Ethyl Acetate	7.559	43	11287	2.006	ug/l	99
43) Isopropyl Acetate	8.688	43	308689	29.829	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020724\
Data File : VN080935.D
Acq On : 07 Feb 2024 14:27
Operator : JC\MD
Sample : P1362-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2

Quant Time: Feb 08 00:25:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 07 06:22:23 2024
Response via : Initial Calibration

