

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020724\
 Data File : VN080938.D
 Acq On : 07 Feb 2024 16:03
 Operator : JC\MD
 Sample : PB158850ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB158850ZHE#02

Quant Time: Feb 08 00:27:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 07 06:22:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	263349	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	711864	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	709944	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	299042	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	192070	57.815	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.620%
35) Dibromofluoromethane	8.165	113	180849	47.344	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.680%
50) Toluene-d8	10.571	98	753805	54.557	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.120%
62) 4-Bromofluorobenzene	12.853	95	295421	58.857	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	117.720%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	34268	31.353	ug/l	100
20) Methylene Chloride	5.277	84	22369	6.755	ug/l #	71
25) 2-Butanone	7.489	43	9837	5.607	ug/l #	79
43) Isopropyl Acetate	8.694	43	234102	29.607	ug/l #	90
95) Naphthalene	15.647	128	31699	1.938	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020724\
Data File : VN080938.D
Acq On : 07 Feb 2024 16:03
Operator : JC\MD
Sample : PB158850ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB158850ZHE#02

Quant Time: Feb 08 00:27:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 07 06:22:23 2024
Response via : Initial Calibration

