

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020724\
 Data File : VN080942.D
 Acq On : 07 Feb 2024 17:39
 Operator : JC\MD
 Sample : PB158850ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB158850ZHE#06

Quant Time: Feb 08 00:29:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 07 06:22:23 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	260749	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	569451	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	711063	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	309038	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	191029	58.075	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 116.140%		
35) Dibromofluoromethane	8.165	113	177094	57.956	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 115.920%		
50) Toluene-d8	10.571	98	694485	62.834	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 125.660%#		
62) 4-Bromofluorobenzene	12.853	95	290451	72.339	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 144.680%#		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	45060	41.638	ug/l	97
20) Methylene Chloride	5.271	84	9381	2.861	ug/l #	82
43) Isopropyl Acetate	8.688	43	295872	46.777	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020724\
Data File : VN080942.D
Acq On : 07 Feb 2024 17:39
Operator : JC\MD
Sample : PB158850ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB158850ZHE#06

Quant Time: Feb 08 00:29:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 07 06:22:23 2024
Response via : Initial Calibration

