

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020724\
 Data File : VN080944.D
 Acq On : 07 Feb 2024 18:27
 Operator : JC\MD
 Sample : PB158850ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB158850ZHE#08

Quant Time: Feb 08 00:30:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 07 06:22:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	276361	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	609234	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	761427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	321564	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	203089	58.253	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.500%
35) Dibromofluoromethane	8.165	113	187252	57.279	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	114.560%
50) Toluene-d8	10.571	98	729233	61.670	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	123.340%#
62) 4-Bromofluorobenzene	12.853	95	310090	72.187	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	144.380%#
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	44622	38.904	ug/l #	80
20) Methylene Chloride	5.277	84	10473	3.014	ug/l #	76
37) Ethyl Acetate	7.565	43	8277	2.249	ug/l	96
43) Isopropyl Acetate	8.688	43	301510	44.556	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020724\
Data File : VN080944.D
Acq On : 07 Feb 2024 18:27
Operator : JC\MD
Sample : PB158850ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB158850ZHE#08

Quant Time: Feb 08 00:30:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 07 06:22:23 2024
Response via : Initial Calibration

