

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020724\
 Data File : VN080945.D
 Acq On : 07 Feb 2024 18:51
 Operator : JC\MD
 Sample : PB158850ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB158850ZHE#01

Quant Time: Feb 08 00:30:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 07 06:22:23 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	252225	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	564084	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	700901	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	299288	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	193463	60.802	ug/l	0.00
Spiked Amount 50.000	Range	74 - 125	Recovery	=	121.600%	
35) Dibromofluoromethane	8.171	113	176335	58.257	ug/l	0.00
Spiked Amount 50.000	Range	75 - 124	Recovery	=	116.520%	
50) Toluene-d8	10.571	98	697017	63.663	ug/l	0.00
Spiked Amount 50.000	Range	86 - 113	Recovery	=	127.320%#	
62) 4-Bromofluorobenzene	12.853	95	293981	73.914	ug/l	0.00
Spiked Amount 50.000	Range	83 - 123	Recovery	=	147.820%#	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	34817	33.260	ug/l	99
20) Methylene Chloride	5.277	84	21105	6.654	ug/l #	83
25) 2-Butanone	7.494	43	10598	6.307	ug/l #	85
37) Ethyl Acetate	7.565	43	10309	3.026	ug/l #	84
43) Isopropyl Acetate	8.688	43	250895	40.044	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020724\
Data File : VN080945.D
Acq On : 07 Feb 2024 18:51
Operator : JC\MD
Sample : PB158850ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB158850ZHE#01

Quant Time: Feb 08 00:30:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 07 06:22:23 2024
Response via : Initial Calibration

