

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020724\
 Data File : VN080928.D
 Acq On : 07 Feb 2024 09:43
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 08 00:20:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 07 06:22:23 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	110	0.00
2 T	Dichlorodifluoromethane	50.000	42.891	14.2	88	0.00
3 P	Chloromethane	50.000	45.372	9.3	95	0.00
4 C	Vinyl Chloride	50.000	43.201	13.6#	97	0.00
5 T	Bromomethane	50.000	41.779	16.4	97	0.00
6 T	Chloroethane	50.000	40.898	18.2	97	0.00
7 T	Trichlorofluoromethane	50.000	44.611	10.8	95	0.00
8 T	Diethyl Ether	50.000	42.440	15.1	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	44.429	11.1	96	0.00
10 T	Methyl Iodide	50.000	46.208	7.6	95	0.00
11 T	Tert butyl alcohol	250.000	205.096	18.0	88	0.00
12 CM	1,1-Dichloroethene	50.000	43.096	13.8#	94	0.00
13 T	Acrolein	250.000	219.139	12.3	89	0.00
14 T	Allyl chloride	50.000	45.673	8.7	100	0.00
15 T	Acrylonitrile	250.000	220.811	11.7	93	0.00
16 T	Acetone	250.000	204.726	18.1	98	0.00
17 T	Carbon Disulfide	50.000	43.434	13.1	97	0.00
18 T	Methyl Acetate	50.000	42.976	14.0	93	0.00
19 T	Methyl tert-butyl Ether	50.000	45.694	8.6	96	0.00
20 T	Methylene Chloride	50.000	42.452	15.1	97	0.00
21 T	trans-1,2-Dichloroethene	50.000	43.066	13.9	96	0.00
22 T	Diisopropyl ether	50.000	47.307	5.4	101	0.00
23 T	Vinyl Acetate	250.000	257.378	-3.0	97	0.00
24 P	1,1-Dichloroethane	50.000	44.756	10.5	99	0.00
25 T	2-Butanone	250.000	205.519	17.8	95	0.00
26 T	2,2-Dichloropropane	50.000	46.762	6.5	99	0.00
27 T	cis-1,2-Dichloroethene	50.000	45.286	9.4	96	0.00
28 T	Bromochloromethane	50.000	50.651	-1.3	109	0.00
29 T	Tetrahydrofuran	250.000	207.438	17.0	91	0.00
30 C	Chloroform	50.000	44.893	10.2#	98	0.00
31 T	Cyclohexane	50.000	41.103	17.8	94	0.00
32 T	1,1,1-Trichloroethane	50.000	46.402	7.2	96	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.466	1.1	97	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	96	0.00
35 S	Dibromofluoromethane	50.000	54.511	-9.0	102	0.00
36 T	1,1-Dichloropropene	50.000	52.759	-5.5	100	0.00
37 T	Ethyl Acetate	50.000	51.019	-2.0	94	0.00
38 T	Carbon Tetrachloride	50.000	53.286	-6.6	96	0.00
39 T	Methylcyclohexane	50.000	44.724	10.6	80	0.00
40 TM	Benzene	50.000	51.734	-3.5	88	0.00
41 T	Methacrylonitrile	50.000	48.070	3.9	92	0.00
42 TM	1,2-Dichloroethane	50.000	50.633	-1.3	91	0.00
43 T	Isopropyl Acetate	50.000	50.068	-0.1	87	0.00
44 TM	Trichloroethene	50.000	44.472	11.1	85	0.00
45 C	1,2-Dichloropropane	50.000	46.517	7.0#	86	0.00
46 T	Dibromomethane	50.000	46.420	7.2	86	0.00
47 T	Bromodichloromethane	50.000	49.062	1.9	91	0.00
48 T	Methyl methacrylate	50.000	40.171	19.7	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020724\
 Data File : VN080928.D
 Acq On : 07 Feb 2024 09:43
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 08 00:20:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 07 06:22:23 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	732.977	26.7#	74	0.00
50 S	Toluene-d8	50.000	53.473	-6.9	102	0.00
51 T	4-Methyl-2-Pentanone	250.000	236.844	5.3	92	0.00
52 CM	Toluene	50.000	49.378	1.2#	98	0.00
53 T	t-1,3-Dichloropropene	50.000	54.097	-8.2	101	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.256	-2.5	98	0.00
55 T	1,1,2-Trichloroethane	50.000	49.487	1.0	96	0.00
56 T	Ethyl methacrylate	50.000	57.670	-15.3	94	0.00
57 T	1,3-Dichloropropane	50.000	51.566	-3.1	100	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	238.906	4.4	83	0.00
59 T	2-Hexanone	250.000	241.271	3.5	95	0.00
60 T	Dibromochloromethane	50.000	53.145	-6.3	96	0.00
61 T	1,2-Dibromoethane	50.000	50.442	-0.9	96	0.00
62 S	4-Bromofluorobenzene	50.000	52.546	-5.1	99	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	108	0.00
64 T	Tetrachloroethene	50.000	42.146	15.7	96	0.00
65 PM	Chlorobenzene	50.000	45.460	9.1	98	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	47.256	5.5	98	0.00
67 C	Ethyl Benzene	50.000	46.748	6.5#	98	0.00
68 T	m/p-Xylenes	100.000	94.517	5.5	98	0.00
69 T	o-Xylene	50.000	48.376	3.2	99	0.00
70 T	Styrene	50.000	53.622	-7.2	98	0.00
71 P	Bromoform	50.000	48.940	2.1	96	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	108	0.00
73 T	Isopropylbenzene	50.000	46.223	7.6	96	0.00
74 T	N-amyl acetate	50.000	44.879	10.2	94	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	42.685	14.6	94	0.00
76 T	1,2,3-Trichloropropane	50.000	44.657	10.7	94	0.00
77 T	Bromobenzene	50.000	44.958	10.1	97	0.00
78 T	n-propylbenzene	50.000	47.487	5.0	98	0.00
79 T	2-Chlorotoluene	50.000	46.338	7.3	98	0.00
80 T	1,3,5-Trimethylbenzene	50.000	47.348	5.3	97	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.553	8.9	84	0.00
82 T	4-Chlorotoluene	50.000	46.247	7.5	95	0.00
83 T	tert-Butylbenzene	50.000	46.048	7.9	95	0.00
84 T	1,2,4-Trimethylbenzene	50.000	47.383	5.2	96	0.00
85 T	sec-Butylbenzene	50.000	48.019	4.0	96	0.00
86 T	p-Isopropyltoluene	50.000	46.937	6.1	94	0.00
87 T	1,3-Dichlorobenzene	50.000	45.314	9.4	96	0.00
88 T	1,4-Dichlorobenzene	50.000	45.534	8.9	99	0.00
89 T	n-Butylbenzene	50.000	46.525	7.0	96	0.00
90 T	Hexachloroethane	50.000	49.953	0.1	101	0.00
91 T	1,2-Dichlorobenzene	50.000	46.275	7.5	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	41.849	16.3	91	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.082	5.8	95	0.00
94 T	Hexachlorobutadiene	50.000	45.328	9.3	97	0.00
95 T	Naphthalene	50.000	46.748	6.5	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020724\
Data File : VN080928.D
Acq On : 07 Feb 2024 09:43
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 08 00:20:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 07 06:22:23 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	45.394	9.2	92	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6