

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
 Data File : VN085694.D
 Acq On : 07 Feb 2025 13:26
 Operator : JC\MD
 Sample : VN0207WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0207WBL01

Quant Time: Feb 07 22:29:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	213382	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	398447	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	349175	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	135868	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	178507	51.825	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.640%
35) Dibromofluoromethane	8.165	113	146516	53.004	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.000%
50) Toluene-d8	10.565	98	489841	49.875	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.760%
62) 4-Bromofluorobenzene	12.847	95	147394	43.872	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
Data File : VN085694.D
Acq On : 07 Feb 2025 13:26
Operator : JC\MD
Sample : VN0207WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0207WBL01

Quant Time: Feb 07 22:29:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

