

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
 Data File : VN085696.D
 Acq On : 07 Feb 2025 14:14
 Operator : JC\MD
 Sample : PB166592TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166592TB

Quant Time: Feb 07 22:30:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	210352	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	410151	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	398261	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	143233	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	178938	52.698	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.400%
35) Dibromofluoromethane	8.165	113	144737	50.866	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.740%
50) Toluene-d8	10.565	98	539986	53.412	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.820%
62) 4-Bromofluorobenzene	12.847	95	160770	46.488	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.980%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	29810	27.171	ug/l	97
43) Isopropyl Acetate	8.689	43	179694	27.821	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
Data File : VN085696.D
Acq On : 07 Feb 2025 14:14
Operator : JC\MD
Sample : PB166592TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166592TB

Quant Time: Feb 07 22:30:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

