

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
 Data File : VN085700.D
 Acq On : 07 Feb 2025 16:41
 Operator : JC\MD
 Sample : PB166592ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166592ZHE#02

Quant Time: Feb 07 22:33:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	205784	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	415637	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	401043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	148113	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	178965	53.876	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.760%	
35) Dibromofluoromethane	8.165	113	144261	50.030	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.060%	
50) Toluene-d8	10.565	98	550011	53.686	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.380%	
62) 4-Bromofluorobenzene	12.847	95	161389	46.051	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 92.100%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	29205	27.211	ug/l	96
43) Isopropyl Acetate	8.688	43	204434	31.234	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
Data File : VN085700.D
Acq On : 07 Feb 2025 16:41
Operator : JC\MD
Sample : PB166592ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166592ZHE#02

Quant Time: Feb 07 22:33:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

