

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
 Data File : VN085704.D
 Acq On : 07 Feb 2025 18:17
 Operator : JC\MD
 Sample : PB166592ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166592ZHE#06

Quant Time: Feb 07 22:34:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	198768	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	405427	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	386691	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	137888	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	173411	54.047	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	108.100%	
35) Dibromofluoromethane	8.165	113	142304	50.594	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.180%	
50) Toluene-d8	10.565	98	528731	52.908	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	105.820%	
62) 4-Bromofluorobenzene	12.847	95	151759	44.394	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	88.780%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	38909	37.531	ug/l	100
18) Methyl Acetate	5.030	43	4037	1.281	ug/l #	86
20) Methylene Chloride	5.283	84	3379	1.317	ug/l #	80
43) Isopropyl Acetate	8.688	43	225725	35.355	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
Data File : VN085704.D
Acq On : 07 Feb 2025 18:17
Operator : JC\MD
Sample : PB166592ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166592ZHE#06

Quant Time: Feb 07 22:34:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

