

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
 Data File : VN085707.D
 Acq On : 07 Feb 2025 19:29
 Operator : JC\MD
 Sample : Q1299-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Feb 07 22:36:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	192305	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	401753	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	379324	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	136382	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	171804	55.346	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	=	110.700%	
35) Dibromofluoromethane	8.165	113	140011	50.234	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	100.460%	
50) Toluene-d8	10.565	98	518577	52.367	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	=	104.740%	
62) 4-Bromofluorobenzene	12.847	95	149428	44.112	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	=	88.220%	
Target Compounds						Qvalue
16) Acetone	4.430	43	26282	26.204	ug/l	93
18) Methyl Acetate	5.030	43	3410	1.118	ug/l #	81
43) Isopropyl Acetate	8.688	43	223336	35.301	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
Data File : VN085707.D
Acq On : 07 Feb 2025 19:29
Operator : JC\MD
Sample : Q1299-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Feb 07 22:36:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

