

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
 Data File : VN085708.D
 Acq On : 07 Feb 2025 19:54
 Operator : JC\MD
 Sample : Q1299-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2

Quant Time: Feb 07 22:36:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.223	168	207399	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	424127	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	400322	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	143708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	182063	54.382	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.760%	
35) Dibromofluoromethane	8.165	113	147581	50.157	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.320%	
50) Toluene-d8	10.565	98	550864	52.693	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.380%	
62) 4-Bromofluorobenzene	12.847	95	157783	44.121	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 88.240%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	50418	46.609	ug/l	99
18) Methyl Acetate	5.018	43	3690	1.122	ug/l	95
20) Methylene Chloride	5.271	84	3409	1.273	ug/l #	79
43) Isopropyl Acetate	8.682	43	215579	32.277	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
Data File : VN085708.D
Acq On : 07 Feb 2025 19:54
Operator : JC\MD
Sample : Q1299-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2

Quant Time: Feb 07 22:36:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

