

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
 Data File : VN085711.D
 Acq On : 07 Feb 2025 21:06
 Operator : JC\MD
 Sample : Q1309-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-4

Quant Time: Feb 07 22:38:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	198721	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	413391	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	388532	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	137402	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	178624	55.685	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.360%
35) Dibromofluoromethane	8.165	113	144508	50.388	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.780%
50) Toluene-d8	10.565	98	541121	53.105	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.200%
62) 4-Bromofluorobenzene	12.847	95	153406	44.011	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.020%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	33155	31.989	ug/l	96
18) Methyl Acetate	5.030	43	4530	1.437	ug/l	96
43) Isopropyl Acetate	8.688	43	247405	38.004	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
Data File : VN085711.D
Acq On : 07 Feb 2025 21:06
Operator : JC\MD
Sample : Q1309-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-4

Quant Time: Feb 07 22:38:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

