

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
 Data File : VN085712.D
 Acq On : 07 Feb 2025 21:30
 Operator : JC\MD
 Sample : Q1309-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-5

Quant Time: Feb 08 00:43:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	202703	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	419706	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	392979	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	140468	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	179366	54.818	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.640%	
35) Dibromofluoromethane	8.165	113	145665	50.027	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.060%	
50) Toluene-d8	10.565	98	540845	52.279	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.560%	
62) 4-Bromofluorobenzene	12.847	95	152381	43.059	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 86.120%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	30670	29.010	ug/l	95
18) Methyl Acetate	5.030	43	3984	1.239	ug/l	91
43) Isopropyl Acetate	8.688	43	238954	36.154	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
Data File : VN085712.D
Acq On : 07 Feb 2025 21:30
Operator : JC\MD
Sample : Q1309-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-5

Quant Time: Feb 08 00:43:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

