

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
 Data File : VN085713.D
 Acq On : 07 Feb 2025 21:54
 Operator : JC\MD
 Sample : Q1309-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-6

Quant Time: Feb 08 00:44:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	196218	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	403083	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	372752	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	130978	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	173046	54.634	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.260%		
35) Dibromofluoromethane	8.165	113	141994	50.777	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.560%		
50) Toluene-d8	10.565	98	527993	53.142	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 106.280%		
62) 4-Bromofluorobenzene	12.847	95	143029	42.083	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 84.160%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	27287	26.663	ug/l	97
18) Methyl Acetate	5.024	43	3935	1.264	ug/l	100
43) Isopropyl Acetate	8.682	43	232789	36.674	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
Data File : VN085713.D
Acq On : 07 Feb 2025 21:54
Operator : JC\MD
Sample : Q1309-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-6

Quant Time: Feb 08 00:44:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

