

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
 Data File : VN085714.D
 Acq On : 07 Feb 2025 22:18
 Operator : JC\MD
 Sample : Q1309-24
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-8

Quant Time: Feb 08 00:44:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	193637	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	407887	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	371092	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	129028	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	174766	55.913	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	111.820%	
35) Dibromofluoromethane	8.159	113	140714	49.727	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.460%	
50) Toluene-d8	10.565	98	525258	52.244	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	104.480%	
62) 4-Bromofluorobenzene	12.847	95	143734	41.793	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	83.580%	
Target Compounds						
						Qvalue
16) Acetone	4.436	43	30765	30.462	ug/l	98
18) Methyl Acetate	5.024	43	4019	1.309	ug/l #	76
43) Isopropyl Acetate	8.688	43	229377	35.711	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020725\
Data File : VN085714.D
Acq On : 07 Feb 2025 22:18
Operator : JC\MD
Sample : Q1309-24
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-8

Quant Time: Feb 08 00:44:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

