

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN020819\
 Data File : VN053682.D
 Acq On : 8 Feb 2019 15:31
 Operator : JC/SP
 Sample : PB116934TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116934TB

Quant Time: Feb 09 03:21:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	866429	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1420500	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1295782	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	545190	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	448341	51.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.16%	
35) Dibromofluoromethane	7.59	113	442741	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
50) Toluene-d8	10.09	98	1724558	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
62) 4-Bromofluorobenzene	12.40	95	583832	51.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.88%	
Target Compounds						
16) Acetone	3.82	43	23095	10.359	ug/l	96
20) Methylene Chloride	4.56	84	15763	1.685	ug/l	94
43) Isopropyl Acetate	8.17	43	40797	2.954	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN020819\
Data File : VN053682.D
Acq On : 8 Feb 2019 15:31
Operator : JC/SP
Sample : PB116934TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB116934TB

Quant Time: Feb 09 03:21:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

