

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020821\
 Data File : VN065749.D
 Acq On : 8 Feb 2021 11:31
 Operator : JC/MD
 Sample : VN0208WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0208WBL01

Quant Time: Feb 09 00:37:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.621	168	186450	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	312101	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	275033	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	95851	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	107627	43.88	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.76%
35) Dibromofluoromethane	7.547	113	90611	48.99	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.98%
50) Toluene-d8	10.058	98	388516	53.44	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.88%
62) 4-Bromofluorobenzene	12.370	95	124015	47.37	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.74%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020821\
Data File : VN065749.D
Acq On : 8 Feb 2021 11:31
Operator : JC/MD
Sample : VN0208WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0208WBL01

Quant Time: Feb 09 00:37:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

