

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020821\
 Data File : VN065752.D
 Acq On : 8 Feb 2021 13:57
 Operator : JC/MD
 Sample : M1292-18
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2021-01-28-EB

Quant Time: Feb 09 00:39:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	194589	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	326733	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	286484	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	102593	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.988	65	113156	44.20	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	88.40%
35) Dibromofluoromethane	7.547	113	93600	48.34	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.68%
50) Toluene-d8	10.058	98	404720	53.18	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.36%
62) 4-Bromofluorobenzene	12.370	95	128536	46.90	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020821\
Data File : VN065752.D
Acq On : 8 Feb 2021 13:57
Operator : JC/MD
Sample : M1292-18
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2021-01-28-EB

Quant Time: Feb 09 00:39:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

