

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020821\
 Data File : VN065762.D
 Acq On : 8 Feb 2021 18:04
 Operator : JC/MD
 Sample : PB134480ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134480ZHE#01

Quant Time: Feb 09 00:42:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	180586	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	302726	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	274393	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	101851	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	103555	43.59	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.18%
35) Dibromofluoromethane	7.547	113	87922	49.01	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.02%
50) Toluene-d8	10.058	98	382015	54.17	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.34%
62) 4-Bromofluorobenzene	12.370	95	124748	49.13	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.26%
Target Compounds						
16) Acetone	3.779	43	13397	14.49	ug/l	99
20) Methylene Chloride	4.496	84	9165	3.58	ug/l	96
43) Isopropyl Acetate	8.126	43	13651	2.63	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020821\
Data File : VN065762.D
Acq On : 8 Feb 2021 18:04
Operator : JC/MD
Sample : PB134480ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134480ZHE#01

Quant Time: Feb 09 00:42:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

