

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020821\
 Data File : VN065763.D
 Acq On : 8 Feb 2021 18:29
 Operator : JC/MD
 Sample : PB134480ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134480ZHE#02

Quant Time: Feb 09 00:42:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	182681	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	307062	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	273037	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	98919	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.988	65	105632	43.95	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.90%
35) Dibromofluoromethane	7.550	113	88494	48.63	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.26%
50) Toluene-d8	10.058	98	379111	53.00	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.00%
62) 4-Bromofluorobenzene	12.373	95	122965	47.74	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.48%
Target Compounds						
16) Acetone	3.782	43	13783	14.73	ug/l	99
20) Methylene Chloride	4.505	84	9519	3.67	ug/l	97
43) Isopropyl Acetate	8.129	43	14201	2.70	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020821\
Data File : VN065763.D
Acq On : 8 Feb 2021 18:29
Operator : JC/MD
Sample : PB134480ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134480ZHE#02

Quant Time: Feb 09 00:42:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

