

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020821\
 Data File : VN065764.D
 Acq On : 8 Feb 2021 18:53
 Operator : JC/MD
 Sample : PB134480ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134480ZHE#03

Quant Time: Feb 09 00:43:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	194694	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	327518	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	299489	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	107788	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	111460	43.52	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	87.04%		
35) Dibromofluoromethane	7.547	113	93166	48.00	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.00%		
50) Toluene-d8	10.058	98	410050	53.75	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	107.50%		
62) 4-Bromofluorobenzene	12.373	95	133760	48.69	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	97.38%		
Target Compounds					Qvalue	
16) Acetone	3.779	43	14072	14.11	ug/l	96
20) Methylene Chloride	4.496	84	10562	3.83	ug/l	98
43) Isopropyl Acetate	8.129	43	13824	2.47	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020821\
Data File : VN065764.D
Acq On : 8 Feb 2021 18:53
Operator : JC/MD
Sample : PB134480ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134480ZHE#03

Quant Time: Feb 09 00:43:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

