

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020821\
 Data File : VN065766.D
 Acq On : 8 Feb 2021 19:43
 Operator : JC/MD
 Sample : PB134480ZHE#05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134480ZHE#05

Quant Time: Feb 09 00:43:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	189324	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	315861	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	284115	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	101610	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	108155	43.42	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	86.84%
35) Dibromofluoromethane	7.547	113	90510	48.36	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.72%
50) Toluene-d8	10.055	98	395912	53.81	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.62%
62) 4-Bromofluorobenzene	12.373	95	127183	48.00	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.00%
Target Compounds						
16) Acetone	3.779	43	13481	13.90	ug/l	100
20) Methylene Chloride	4.496	84	10152	3.78	ug/l	94
43) Isopropyl Acetate	8.129	43	14049	2.60	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020821\
Data File : VN065766.D
Acq On : 8 Feb 2021 19:43
Operator : JC/MD
Sample : PB134480ZHE#05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134480ZHE#05

Quant Time: Feb 09 00:43:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

