

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020821\
 Data File : VN065769.D
 Acq On : 8 Feb 2021 20:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 09 00:44:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.560	0.522	6.8	74	0.00
3 P	Chloromethane	0.823	0.769	6.6	74	0.00
4 C	Vinyl Chloride	0.833	0.790	5.2#	76	0.00
5 T	Bromomethane	0.482	0.469	2.7	77	-0.02
6 T	Chloroethane	0.482	0.458	5.0	74	0.00
7 T	Trichlorofluoromethane	0.871	0.871	0.0	79	0.00
8 T	Diethyl Ether	0.415	0.386	7.0	74	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.529	0.528	0.2	81	0.00
10 T	Methyl Iodide	0.775	0.840	-8.4	86	0.00
11 T	Tert butyl alcohol	0.123	0.100	18.7	65	0.00
12 CM	1,1-Dichloroethene	0.609	0.553	9.2#	76	0.00
13 T	Acrolein	0.095	0.109	-14.7	83	0.00
14 T	Allyl chloride	1.128	0.964	14.5	67	0.00
15 T	Acrylonitrile	0.321	0.291	9.3	71	0.00
16 T	Acetone	0.256	0.212	17.2	67	0.00
17 T	Carbon Disulfide	1.916	1.447	24.5	65	0.00
18 T	Methyl Acetate	0.718	0.690	3.9	79	0.00
19 T	Methyl tert-butyl Ether	2.032	1.828	10.0	73	0.00
20 T	Methylene Chloride	0.709	0.657	7.3	78	0.00
21 T	trans-1,2-Dichloroethene	0.647	0.585	9.6	75	0.00
22 T	Diisopropyl ether	2.307	2.016	12.6	70	0.00
23 T	Vinyl Acetate	1.842	1.562	15.2	66	0.00
24 P	1,1-Dichloroethane	1.229	1.129	8.1	72	0.00
25 T	2-Butanone	0.402	0.354	11.9	68	0.00
26 T	2,2-Dichloropropane	0.984	0.787	20.0	65	0.00
27 T	cis-1,2-Dichloroethene	0.741	0.697	5.9	77	0.00
28 T	Bromochloromethane	0.517	0.417	19.3	69	0.00
29 T	Tetrahydrofuran	0.291	0.241	17.2	68	0.00
30 C	Chloroform	1.132	1.061	6.3#	75	0.00
31 T	Cyclohexane	1.237	0.993	19.7	67	0.00
32 T	1,1,1-Trichloroethane	0.922	0.860	6.7	75	0.00
33 S	1,2-Dichloroethane-d4	0.658	0.578	12.2	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.296	0.288	2.7	80	0.00
36 T	1,1-Dichloropropene	0.503	0.466	7.4	73	0.00
37 T	Ethyl Acetate	0.483	0.434	10.1	68	0.00
38 T	Carbon Tetrachloride	0.414	0.395	4.6	74	0.00
39 T	Methylcyclohexane	0.614	0.530	13.7	68	0.00
40 TM	Benzene	1.537	1.497	2.6	76	0.00
41 T	Methacrylonitrile	0.259	0.216	16.6	67	0.00
42 TM	1,2-Dichloroethane	0.463	0.446	3.7	74	0.00
43 T	Isopropyl Acetate	0.856	0.719	16.0	67	0.00
44 TM	Trichloroethene	0.373	0.372	0.3	79	0.00
45 C	1,2-Dichloropropane	0.423	0.406	4.0#	74	0.00
46 T	Dibromomethane	0.234	0.229	2.1	75	0.00
47 T	Bromodichloromethane	0.501	0.471	6.0	73	0.00
48 T	Methyl methacrylate	0.406	0.349	14.0	67	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020821\
 Data File : VN065769.D
 Acq On : 8 Feb 2021 20:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 09 00:44:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.006	14.3	75	0.00
50 S	Toluene-d8	1.165	1.121	3.8	80	0.00
51 T	4-Methyl-2-Pentanone	0.472	0.439	7.0	71	0.00
52 CM	Toluene	0.910	0.915	-0.5#	79	0.00
53 T	t-1,3-Dichloropropene	0.583	0.548	6.0	72	0.00
54 T	cis-1,3-Dichloropropene	0.650	0.621	4.5	74	0.00
55 T	1,1,2-Trichloroethane	0.347	0.365	-5.2	82	0.00
56 T	Ethyl methacrylate	0.574	0.555	3.3	74	0.00
57 T	1,3-Dichloropropane	0.611	0.618	-1.1	78	0.00
58 T	2-Chloroethyl Vinyl ether	0.265	0.237	10.6	69	0.00
59 T	2-Hexanone	0.328	0.309	5.8	71	0.00
60 T	Dibromochloromethane	0.355	0.362	-2.0	79	0.00
61 T	1,2-Dibromoethane	0.346	0.359	-3.8	80	0.00
62 S	4-Bromofluorobenzene	0.419	0.425	-1.4	85	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64 T	Tetrachloroethene	0.419	0.379	9.5	78	0.00
65 PM	Chlorobenzene	1.046	1.056	-1.0	84	0.00
66 T	1,1,1,2-Tetrachloroethane	0.370	0.368	0.5	82	0.00
67 C	Ethyl Benzene	1.965	1.860	5.3#	79	0.00
68 T	m/p-Xylenes	0.711	0.709	0.3	82	0.00
69 T	o-Xylene	0.689	0.700	-1.6	85	0.00
70 T	Styrene	1.135	1.165	-2.6	83	0.00
71 P	Bromoform	0.256	0.250	2.3	78	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	4.378	4.047	7.6	82	0.00
74 T	N-amyl acetate	1.852	1.535	17.1	70	0.00
75 P	1,1,2,2-Tetrachloroethane	1.293	1.215	6.0	85	0.00
76 T	1,2,3-Trichloropropane	1.307	1.139	12.9	79	0.00
77 T	Bromobenzene	0.948	0.954	-0.6	90	0.00
78 T	n-propylbenzene	5.159	4.562	11.6	79	0.00
79 T	2-Chlorotoluene	3.064	2.754	10.1	81	0.00
80 T	1,3,5-Trimethylbenzene	3.727	3.425	8.1	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.510	0.407	20.2	70	0.00
82 T	4-Chlorotoluene	3.128	2.794	10.7	80	0.00
83 T	tert-Butylbenzene	3.108	2.870	7.7	84	0.00
84 T	1,2,4-Trimethylbenzene	3.718	3.419	8.0	83	0.00
85 T	sec-Butylbenzene	4.196	3.781	9.9	81	0.00
86 T	p-Isopropyltoluene	3.705	3.343	9.8	81	0.00
87 T	1,3-Dichlorobenzene	1.690	1.668	1.3	87	0.00
88 T	1,4-Dichlorobenzene	1.715	1.656	3.4	87	0.00
89 T	n-Butylbenzene	3.329	2.804	15.8	74	0.00
90 T	Hexachloroethane	0.648	0.572	11.7	77	0.00
91 T	1,2-Dichlorobenzene	1.676	1.647	1.7	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.211	17.9	70	0.00
93 T	1,2,4-Trichlorobenzene	0.891	0.835	6.3	76	0.00
94 T	Hexachlorobutadiene	0.461	0.358	22.3	77	0.00
95 T	Naphthalene	2.867	2.650	7.6	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020821\
Data File : VN065769.D
Acq On : 8 Feb 2021 20:57
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 09 00:44:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.922	0.849	7.9	79	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6