

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
 Data File : VN076568.D
 Acq On : 08 Feb 2023 10:21
 Operator : JC\MD
 Sample : VN0208MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0208MBL01

Quant Time: Feb 09 00:41:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 09:04:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	541019	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	888493	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	771620	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	303301	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	299204	46.856	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.720%
35) Dibromofluoromethane	8.177	113	267960	48.516	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.040%
50) Toluene-d8	10.571	98	1019340	49.058	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.120%
62) 4-Bromofluorobenzene	12.853	95	340166	46.470	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
Data File : VN076568.D
Acq On : 08 Feb 2023 10:21
Operator : JC\MD
Sample : VN0208MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0208MBL01

Quant Time: Feb 09 00:41:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 09:04:43 2023
Response via : Initial Calibration

