

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
 Data File : VN076569.D
 Acq On : 08 Feb 2023 10:45
 Operator : JC\MD
 Sample : VN0208WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0208WBL01

Quant Time: Feb 09 00:42:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 09:04:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	646060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	1063827	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	910658	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	353694	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	360299	47.250	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.500%
35) Dibromofluoromethane	8.177	113	316947	47.927	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.860%
50) Toluene-d8	10.571	98	1215972	48.876	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.760%
62) 4-Bromofluorobenzene	12.853	95	401580	45.818	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
Data File : VN076569.D
Acq On : 08 Feb 2023 10:45
Operator : JC\MD
Sample : VN0208WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0208WBL01

Quant Time: Feb 09 00:42:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 09:04:43 2023
Response via : Initial Calibration

