

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
 Data File : VN076572.D
 Acq On : 08 Feb 2023 12:10
 Operator : JC\MD
 Sample : PB150715TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150715TB

Quant Time: Feb 09 00:44:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 09:04:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	617062	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	1020802	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	916627	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	364325	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	340926	46.810	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.620%		
35) Dibromofluoromethane	8.171	113	305762	48.185	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.360%		
50) Toluene-d8	10.571	98	1201610	50.335	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	100.660%		
62) 4-Bromofluorobenzene	12.853	95	415001	49.345	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.700%		
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	100436	41.828	ug/l	95
20) Methylene Chloride	5.289	84	37535	5.071	ug/l	96
37) Ethyl Acetate	7.571	43	64015	9.431	ug/l	96
43) Isopropyl Acetate	8.694	43	317272	28.090	ug/l #	92
95) Naphthalene	15.641	128	23308	1.388	ug/l	100
96) 1,2,3-Trichlorobenzene	15.847	180	5908	1.069	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
Data File : VN076572.D
Acq On : 08 Feb 2023 12:10
Operator : JC\MD
Sample : PB150715TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150715TB

Quant Time: Feb 09 00:44:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 09:04:43 2023
Response via : Initial Calibration

