

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
 Data File : VN076574.D
 Acq On : 08 Feb 2023 12:58
 Operator : JC\MD
 Sample : 01418-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JPP-18A-020223

Quant Time: Feb 09 00:45:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 09:04:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	490842	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	765299	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	721319	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	337550	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	227175	39.213	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	78.420%
35) Dibromofluoromethane	8.177	113	222577	46.786	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.580%
50) Toluene-d8	10.570	98	885646	49.485	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.980%
62) 4-Bromofluorobenzene	12.853	95	326034	51.709	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.420%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	24524	12.840	ug/l	96
20) Methylene Chloride	5.300	84	19598	3.328	ug/l #	75
37) Ethyl Acetate	7.571	43	39762	7.814	ug/l #	92
43) Isopropyl Acetate	8.694	43	204337	24.131	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
Data File : VN076574.D
Acq On : 08 Feb 2023 12:58
Operator : JC\MD
Sample : 01418-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-18A-020223

Quant Time: Feb 09 00:45:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 09:04:43 2023
Response via : Initial Calibration

