

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
 Data File : VN076578.D
 Acq On : 08 Feb 2023 14:33
 Operator : JC\MD
 Sample : 01462-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TCLP-2

Quant Time: Feb 09 00:48:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 09:04:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.234	168	605279	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.104	114	1003788	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.869	117	897851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.798	152	356758	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.586	65	330997	46.331	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.660%
35) Dibromofluoromethane	8.175	113	299642	48.021	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.040%
50) Toluene-d8	10.569	98	1185727	50.511	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.020%
62) 4-Bromofluorobenzene	12.851	95	396146	47.902	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.800%
Target Compounds						
					Qvalue	
16) Acetone	4.440	43	70229	29.817	ug/l	99
20) Methylene Chloride	5.293	84	33149	4.565	ug/l	94
37) Ethyl Acetate	7.569	43	63650	9.536	ug/l	98
43) Isopropyl Acetate	8.698	43	343650	30.941	ug/l #	86
95) Naphthalene	15.645	128	18980	1.155	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
Data File : VN076578.D
Acq On : 08 Feb 2023 14:33
Operator : JC\MD
Sample : 01462-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TCLP-2

Quant Time: Feb 09 00:48:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 09:04:43 2023
Response via : Initial Calibration

