

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
 Data File : VN076581.D
 Acq On : 08 Feb 2023 15:46
 Operator : JC\MD
 Sample : 01465-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TCLP-1

Quant Time: Feb 09 00:49:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 09:04:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.233	168	548257	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.110	114	862075	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.869	117	749438	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.792	152	331542	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.586	65	258585	39.960	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	79.920%
35) Dibromofluoromethane	8.175	113	248343	46.342	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.680%
50) Toluene-d8	10.569	98	989123	49.063	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.120%
62) 4-Bromofluorobenzene	12.851	95	340382	47.925	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.840%
Target Compounds						
					Qvalue	
16) Acetone	4.451	43	76476	35.846	ug/l	92
20) Methylene Chloride	5.293	84	17726	2.695	ug/l #	87
27) cis-1,2-Dichloroethene	7.498	96	8249	1.184	ug/l	92
37) Ethyl Acetate	7.575	43	40809	7.119	ug/l #	91
43) Isopropyl Acetate	8.698	43	245401	25.727	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
Data File : VN076581.D
Acq On : 08 Feb 2023 15:46
Operator : JC\MD
Sample : 01465-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TCLP-1

Quant Time: Feb 09 00:49:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 09:04:43 2023
Response via : Initial Calibration

