

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
 Data File : VN076582.D
 Acq On : 08 Feb 2023 16:10
 Operator : JC\MD
 Sample : 01418-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JPP-16-020223

Quant Time: Feb 09 00:50:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 09:04:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.233	168	512497	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.104	114	836086	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.868	117	799032	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.792	152	326271	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.586	65	241076	39.854	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	79.700%
35) Dibromofluoromethane	8.175	113	235297	45.272	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.540%
50) Toluene-d8	10.569	98	1016044	51.964	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.920%
62) 4-Bromofluorobenzene	12.851	95	393925	57.187	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.380%
Target Compounds						
					Qvalue	
16) Acetone	4.445	43	49530	24.836	ug/l	97
20) Methylene Chloride	5.298	84	17731	2.884	ug/l #	86
37) Ethyl Acetate	7.575	43	39427	7.092	ug/l #	91
43) Isopropyl Acetate	8.692	43	238816	25.815	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
Data File : VN076582.D
Acq On : 08 Feb 2023 16:10
Operator : JC\MD
Sample : 01418-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-16-020223

Quant Time: Feb 09 00:50:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 09:04:43 2023
Response via : Initial Calibration

